	Application No.	Applicant(s)
Notice of Allowability		
	10/729,599 Examiner	JASSAWALLA ET AL.  Art Unit
-		2700
	George Manuel	3762
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to		
2.  The allowed claim(s) is/are <u>1-29,31-33 and 36-40</u> .		
3. The drawings filed on <u>05 December 2003</u> are accepted by the Examiner.		
<ul> <li>4. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).</li> <li>a) All b) Some* c) None of the: <ol> <li>Certified copies of the priority documents have been received.</li> <li>Certified copies of the priority documents have been received in Application No.</li> <li>Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).</li> </ol> </li> <li>* Certified copies not received:</li> </ul>		
Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		
5. A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.		
<ul> <li>6. CORRECTED DRAWINGS ( as "replacement sheets") must be submitted.</li> <li>(a) including changes required by the Notice of Draftsperson's Patent Drawing Review ( PTO-948) attached</li> <li>1) hereto or 2) to Paper No./Mail Date</li> <li>(b) including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date</li> <li>Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).</li> </ul>		
7. DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.		
Attachment(s)  1. Notice of References Cited (PTO-892)  2. Notice of Draftperson's Patent Drawing Review (PTO-948)  3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date	6. ⊠ Interview Summary Paper No./Mail Da 08), 7. ⊠ Examiner's Amend	ite
		George Manuel Primary Examine:

Application/Control Number: 10/729,599

Art Unit: 3762

## **DETAILED ACTION**

## Examiner's Amendment

An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.31 2. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Donald L. Bartels on 5/26/05.

The application has been amended as follows:

Cancel claims 34 and 35.

## Allowable Subject Matter

Claims 1-29 and 31-33 and 36-40 are allowed.

The following is an examiner's statement of reasons for allowance:

A ventricular assist device comprising a drive unit adapted to alternately move an actuator towards one or the other of a pair of compressible chambers where the movement of the actuator towards the first chamber is a transfer stroke that transfers the blood within the first chamber to the second chamber as claimed is not taught nor suggested by the prior art of record. Also, a sensor that detects a negative pressure within a first chamber of a ventricular assist device comprising a pair of compressible chambers is not an obvious adaptation over the prior art of record. Further, the sum of the spring forces and the magnetic force providing a net bias force that is approximately independent of the position of the armature along the polar axis and biases the

Art Unit: 3762

armature towards one of the pair of poles in an electromagnetic drive is not rendered obvious from the prior art of record.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

## Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George Manuel whose telephone number is (703) 308-2118.

George Manuel Inmary Examiner Art Unit: 3762

5/26/05